



Presentation shows to similar calibration kit.

Contents

Device	Part number	Quantity	Calibration Option ^a
Open circuit plug	P5S12L-00AS3	1	FC / AC
Open circuit jack	P5K12L-00AS3	1	FC / AC
Short circuit plug	P5S12S-00AS3	1	FC / AC
Short circuit jack	P5K12S-00AS3	1	FC / AC
Calibration load plug	P5S170-C1AS3	1	FC / AC
Calibration load jack	P5K170-C1AS3	1	FC / AC
Calibration Adaptor RPC-N 75 Ω plug / BNC 75 Ω plug	P5S171-S2AS3	2	FC
Calibration Adaptor RPC-N 75 Ω plug / BNC 75 Ω jack	P5S171-K2AS3	2	FC
Combi wrench	53W011-000	1	-
Torque wrench	53W009-000	1	FC

a. See "Declaration of calibration options" for explanation.

Documentation

This kit is delivered with

- **USB-Stick**
Standard Definitions as data files for Vector Network Analyzer Families PNA (Keysight/Agilent) and ZVA (Rohde&Schwarz). Calibration Certificate as PDF-file.
- **Standard Definitions Cards**
Printed Standard Definitions that can be used on nearly all Vector Network Analyzers.
- **Kit Info Card**
Handling precautions and information for installing Standard Definitions on a Vector Network Analyzer.
- **Calibration Certificate**
Details see "Declaration of calibration options"
- **Operating Manual**

Electrical specifications

This specification covers electrical key values for the main calibration standards of the calibration kit. Specific datasheets are available for each component among the part number.

Calibration standard	Frequency	Parameter	Specification
Opens^b (plug and jack)	DC to ≤ 4 GHz > 4 GHz to ≤ 8 GHz > 8 GHz to ≤ 12 GHz	Error from Nominal Phase	≤ 1.5° ≤ 3.0° ≤ 4.0°
Shorts^b (plug and jack)	DC to ≤ 4 GHz > 4 GHz to ≤ 8 GHz > 8 GHz to ≤ 12 GHz	Error from Nominal Phase	≤ 1.5° ≤ 3.0° ≤ 4.0°
Calibration loads (plug and jack)	DC to ≤ 4 GHz > 4 GHz to ≤ 8 GHz > 8 GHz to ≤ 12 GHz	Return Loss	≥ 40 dB ≥ 32 dB ≥ 30 dB
Calibration adaptors (plug/plug and plug/jack)	DC to ≤ 4 GHz > 4 GHz to ≤ 8 GHz > 8 GHz to ≤ 12 GHz	Return Loss	≥ 30 dB ≥ 25 dB ≥ 15 dB

b. The specifications for opens and shorts are given as allowed deviation from nominal model as defined in calibration certificate included with your kit.

Declaration of calibration options

Factory Calibration

Standard delivery for this kit includes a Factory Calibration. All devices marked with "FC" in the Content table above are reported in a Calibration Certificate with their individual calibration results, traceable to national / international standards. Model based definitions of the calibration standards are reported in Agilent/Keysight, Rohde & Schwarz and Anritsu compatible VNA format.

Accredited Calibration

Optional this kit can be delivered with an Accredited Calibration (DAkkS) having the highest confidence in the traceability. All devices marked with "AC" in the Content table above are reported in a DAkkS Calibration Certificate with their individual calibration results in a complex format, traceable to national / international standards. Model based definitions of the calibration standards are reported in Agilent/Keysight, Rohde & Schwarz and Anritsu VNA format as well as in dense data sets needed for data based calibration kits. The uncertainties are a little bit smaller than in a Factory Calibration.

All devices marked with "FC" only cannot be calibrated under accreditation. They are factory calibrated as described above.

For further, more detailed information see application note AN001 on the Rosenberger homepage.

Calibration interval

Recommendation 12 months

Recommended accessories

- Rosenberger Test Port Adaptor
- Rosenberger Gauge Kit 05GK0KS-010
- Rosenberger VNA Test cable kit and Microwave Cable Assemblies

For further, more detailed information please visit our homepage www.rosenberger.com.

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

Draft	Date	Approved	Date	Rev.	Engineering change number	Name	Date
Herbert Babinger	23.09.14	Martin Moder	01.08.16	c00	16-1267	M. Knoll	01.08.16
Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de					Tel. : +49 8684 18-0 Fax : +49 8684 18-499 Email : info@rosenberger.de		Page 3 / 3